


Microscale Acid-Base Titration

A titration is a process used to determine the volume of a solution needed to react with a given amount of another substance. In this experiment, you will titrate hydrochloric acid solution, HCl, with a basic sodium hydroxide solution, NaOH. The concentration of the NaOH solution is given and you will determine the unknown concentration of the HCl. Hydrogen ions from the HCl react with hydroxide ions from the NaOH in a one-to-one ratio to produce water in the overall reaction:



When HCl solution is titrated with NaOH solution, the pH value of the acidic solution is initially low. As base is added, the change in pH is quite gradual until close to the equivalence point, when equimolar amounts of acid and base have been mixed. Near the equivalence point, the pH increases very rapidly. The change in pH then becomes more gradual again, before leveling off with the addition of excess base.

Since this experiment may be your introduction to acid-base titrations, you will determine only the *approximate* concentration of the hydrochloric acid solution. Use the formula:

$$M_{\text{acid}} = M_{\text{base}} \times \frac{V_{\text{base}}}{V_{\text{acid}}}$$

where M_{acid} is the concentration of the acid (in M or mol/L), M_{base} is the concentration of the base, V_{base} is the volume of the base (in drops), and V_{acid} is the volume of the acid. The concentration of the sodium hydroxide solution is 0.10 M. The drops of sodium hydroxide and hydrochloric acid solutions at the equivalence point will be determined from the experiment.


MATERIALS

TI-83 Plus or TI-84 Plus graphing calculator	0.10 M NaOH solution (in dropper bottle)
EasyData application	HCl solution (in a dropper bottle)
data-collection interface	ring stand
Vernier pH Sensor	utility clamp
wash bottle	phenolphthalein indicator
distilled water	micro-beaker (top half of a storage bottle
toothpick (for stirring)	for the pH Sensor)

PROCEDURE

1. Obtain and wear goggles.
2. Prepare the pH Sensor for data collection.
 - a. Turn on the calculator. Connect the pH Sensor, data-collection interface, and calculator.
 - b. Remove the pH Sensor from the pH storage solution bottle by unscrewing the lid. Carefully slide the lid from the sensor body.
 - c. Rinse the tip of the sensor with distilled water.

Experiment 36

- d. Obtain a pH Sensor storage bottle that has been cut in half. This is your microbeaker!
 - e. With the open end of the pH Sensor pointing upward, as shown here, slip the microbeaker and cap down onto the sensor body (small opening first), so the sensor tip extends about 1 cm into the bowl of the microbeaker. Then tighten the threads of the cap so the cap tightens snugly against the pH Sensor body.
 - f. Attach the utility clamp to a ring stand and to the bottle lid, with the sensor in an inverted position as shown here.
- 
3. Obtain a dropper bottle containing the HCl solution of unknown concentration. Add 10 drops of the HCl solution into the micro-beaker. As you add the drops, hold the bottle in a vertical position to ensure that drop size is uniform. **CAUTION:** *Handle the hydrochloric acid with care. It can cause painful burns if it comes in contact with the skin.* Add 1 drop of phenolphthalein indicator to the microbeaker, then add enough distilled water so the resulting solution completely covers the sensor tip. Stir the solution thoroughly with the toothpick.
 4. Obtain a dropper bottle containing 0.10 M NaOH. Wait until Step 6 to begin adding this solution to the HCl solution in the microbeaker.
 5. Set up the calculator and interface for data collection.
 - a. Start the EasyData application, if it is not already running.
 - b. Select **[File]** from the Main screen, and then select **New** to reset the application.
 - c. Select **[Setup]** from the Main screen, and then select **Events with Entry**.
 6. You are now ready to perform the titration. This process goes faster if one person adds drops of NaOH solution while another person operates the calculator and enters volumes.
 - a. Select **[Start]** to begin data collection.
 - b. Before you have added any drops of NaOH solution, select **[Keep]** and enter **0** as the NaOH volume, in drops. Select **[OK]** to save the first data pair for this experiment.
 - c. Add one drop of NaOH solution. Be sure to hold the dropper bottle vertically to ensure that the drop size is uniform. **CAUTION:** *Sodium hydroxide solution is caustic. Avoid spilling it on your skin or clothing.* Stir with a toothpick to uniformly mix the solution. When the pH stabilizes, select **[Keep]**, enter **1** as the number of drops of NaOH solution added, and select **[OK]**. You have now saved the second data pair for the experiment.
 - d. Add a second drop of NaOH solution and stir. When the pH stabilizes, select **[Keep]**, enter **2** as the number of drops added, and select **[OK]**.
 - e. Continue this procedure until 20 drops of NaOH solution have been added.
 7. Select **[Stop]** when you have finished collecting data.
 8. Examine the data on the displayed graph to find the *equivalence point*—that is, the 1 drop volume increment that resulted in the largest increase in pH. As you move the cursor right or left on the displayed graph, the volume (X) and pH (Y) values of each data point are displayed above the graph. Go to the region of the graph with the large increase in pH. Find the NaOH volume (in drops) just *before* this jump. Record this value in the data table. Then record the NaOH volume *after* the drop producing the largest pH increase was added.
 9. (optional) Print a copy of the graph of pH vs. volume.

- Carefully loosen the pH Sensor from the utility clamp and dispose of the beaker contents as directed by your instructor. Thoroughly rinse out the microbeaker and the pH Sensor tip. Remove the microbeaker from the sensor. Rinse the sensor tip with distilled water and return it to the pH Sensor storage bottle.

PROCESSING THE DATA

- Use your printed graph to confirm the volume of NaOH titrant you recorded *before* and *after* the largest increase in pH values upon the addition of 1 drop of NaOH solution.
- Determine the volume of NaOH added at the equivalence point. To do this, add the two NaOH values determined above and divide by two (use 0.5 drop increments in your answer).
- Using the formula in the introduction of the experiment, calculate the concentration of the hydrochloric acid solution (in M or mol/L).

DATA AND CALCULATIONS TABLE

Concentration of NaOH	M
NaOH volume added <i>before</i> the largest pH increase	drops
NaOH volume added <i>after</i> the largest pH increase	drops
Volume of NaOH added at equivalence point	drops
Concentration of HCl	M